

CERTIFICATE of CALIBRATION

for LI-COR SENSOR

Model Number: LI-200SZ

Serial Number: PY55553

Calibration Date: August 17, 2006

Output: 102.2 microamps per 1000 watts m⁻²

IMPORTANT: *Read the appropriate instruction manual before using this sensor.*

IMPORTANT: *It is recommended that sensors be recalibrated every two years.*

Calibration Technician:

Caron Deschane

LI-COR[®]

Biosciences

LI-COR Biosciences • Environmental • 4421 Superior Street • P.O. Box 4425 • Lincoln, NE 68504
Phone: 402-467-3576 • FAX: 402-467-2819 • Toll-free 1-800-447-3576 (U.S. & Canada)
envsales@licor.com • envsupport@licor.com • www.licor.com