

CERTIFICATE OF CALIBRATION

Model Number: **LI-200SZ PYRANOMETER SENSOR**

Serial Number: PY57066

Calibration Constant: 98.2

Calibration Multiplier: -10.18

Units: microamps per 1000 watts m⁻²

Units: watts m⁻² per microamp

Please consult the instruction manual for further information on the calibration constant and calibration multiplier. Recalibration is recommended every two years.

Date of Calibration: May 9, 2007

By: Caron Deschane

LI-COR[®]

Biosciences

4421 Superior Street ● P.O. Box 4425 ● Lincoln, Nebraska 68504 USA

Phone: 402-467-3576 ● FAX: 402-467-2819

Toll-free 1-800-447-3576 (U.S. & Canada)

E-mail: envsales@licor.com

www.licor.com