CERTIFICATE OF CALIBRATION

Model Number: **LI-200SZ PYRANOMETER SENSOR**

Serial Number: PY57534

Calibration Constant: 95.8

Calibration Multiplier: -10.44

Units: microamps per 1000 watts m^{-2}

Units: watts m^{-2} per microamp

Please consult the instruction manual for further information on the calibration constant and calibration multiplier. Recalibration is recommended every two years.

Date of Calibration: June 8, 2007

By: Caron Deschane

LI-COR®
Biosciences

4421 Superior Street ● P.O. Box 4425 ● Lincoln, Nebraska 68504 USA
Phone: 402-467-3576 ● FAX: 402-467-2819
Toll-free 1-800-447-3576 (U.S. & Canada)
E-mail: envsales@licor.com
www.licor.com