



CRBasic

IRR-P 1293 Custom Coefficients

Calibration Date: 13-Mar-2007

	C2	C1	C0
m	65775.4	9644660	1575440000
b	27463	-643138	-5059330

Use these coefficients under "Declare Constants" section in the example Campbell Scientific datalogger program



Edlog

IRR-P 1293 Custom Coefficients

Calibration Date: 13-Mar-2007

	C2	C1	C0
m(SB)	0.65776	96.44756	15754.55755
b(SB)	0.27463	-6.43144	-50.59381

Use these coefficients in instructions 11&13 in the example Campbell Scientific datalogger program