



CRBasic

IRR-P 1299 Custom Coefficients

Calibration Date: 15-Mar-2007

|   | C2      | C1      | C0         |
|---|---------|---------|------------|
| m | 80687.2 | 8752140 | 1475510000 |
| b | 16754.9 | -112532 | 4990630    |

Use these coefficients under "Declare Constants" section in the example Campbell Scientific datalogger program



Edlog

IRR-P 1299 Custom Coefficients

Calibration Date: 15-Mar-2007

|       | C2      | C1       | C0          |
|-------|---------|----------|-------------|
| m(SB) | 0.80688 | 87.52228 | 14755.24755 |
| b(SB) | 0.16755 | -1.12533 | 49.90680    |